

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 08 2009

3 NOV 11 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130177</u>	
Name <u>STEVEN C KEVAN</u>		ID# <u>6585</u>	Date <u>8/25/09</u>
A	Agency <u>ALASKA STATE TROOPERS</u>		Phone # <u>800-240-2019</u>
Instrument Location <u>ST. MARYS</u> ✓			
Alco S/N <u>X 301550</u> ✓		Target Value <u>.079</u>	High Pressure <u>550</u>
B	Alco Test Value Average <u>.079</u> ✓		
	1 st Alco	<u>.080</u> ✓	2 nd Alco
Signature <u>[Signature]</u>			
(OVER)			

BMB 11/11/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 30th day of Dec., 2009.

[Signature]
(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130177 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177 ✓

OCTOBER 25, 2009 ✓

OPERATOR'S NAME:

KEVIN, STEVE ✓

OPERATOR'S NUMBER: 6585

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

T

O.L. #: 00

DEPT/AGENCY: SMO

CASE/REPORT: 0

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X301550 ✓

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.49 in
ALCO TARGET .077 13:52
BLANK TEST .000 13:53
INTERNAL STANDARD VERIFIED 13:53
ALCO TV 29.49 in .079 ✓ 13:53
BLANK TEST .000 13:54
SUBJECT SAMPLE .000 13:55
BLANK TEST .000 13:55
ALCO TV 29.51 in .000 ✓ 13:56
BLANK TEST .000 13:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130177 ✓

OCTOBER 25, 2009 ✓

TIME 13:49

— DIAGNOSTIC CHECK —

COMPUTER: OKAY ✓

PROGRAM: OKAY ✓

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 29.49 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEF6
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~
printuvxyz(1)~